

***National Type Evaluation Program  
Certificate of Conformance  
for Weighing and Measuring Devices***

**For:**

Indicating Element  
Digital Electronic  
Model: AD 4323  
n: 10,000

**Submitted by:**


A&D Engineering  
1555 McCandless Drive  
Milpitas, CA 95035  
Contact: Steve Lipham  
Tel.: (408) 263-5333  
Fax: (408) 263-0119

**Standard Features and Options**

Semi-automatic zero  
Semi-automatic tare  
Automatic zero setting mechanism  
Gross/net display modes  
Pound/kilogram conversion  
Printer interface  
Keyboard calibration

This device was evaluated under the National Type Evaluation Program (NTEP) and was found to comply with the applicable technical requirements of Handbook 44, "Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Effective Date: June 8, 1988



Louis E. Straub  
Chairman, NCWM, Inc.



G. Weston Diggs  
Chairman, National Type Evaluation Program Committee

Issue date: June 8, 1988

Note: The National Conference on Weights and Measures does not "approve", "recommend", or "endorse" any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.

This is a reissuance by the NCWM of a Certificate of Conformance already issued by the National Institute of Standards and Technology.

**A And D Engineering  
Indicating Element  
Model: AD 4323**

**Application:**

General.

**Sealing:**

Access to the calibration mode is sealable using a lead and wire seal on the front of the indicator.

**Test Condition:**

The emphasis of the examination was on the device design, operation and compliance for influence factor requirements for accuracy. The indicating element was interfaced to a load cell simulator and tested over a temperature range of -10 °C to 40 °C. Additionally, the device was tested at 100, 120 and 130 v. The results indicate the device complies with the applicable requirements of NBS Handbook 44.

**Type Evaluation Criteria Used:**

1987 Edition

**Tested By:**

Steve Cook and Sally Barron