

*National Type Evaluation Program
Certificate of Conformance
for Weighing and Measuring Devices*

For:

Bench and Counter Scale
Digital Electronic
Models: HV-XXKGV*, HV-XXKGL*,
and HV-XXKV-WP*
 n_{\max} : 3000
Capacities: 15 kg (30 lb), 60 kg (150 lb),
and 220 kg (500 lb)
Accuracy Class: III

Submitted by:

A&D Engineering
1555 McCandless Drive
Milpitas, CA 95035
Tel: (408) 263-5333
Fax: (408) 263-0119
Contact: Jesus Zapien

Standard Features and Options

* See Page 2 for specific model designations and e_{\min} values.

Automatic multiple range operation

Gross/net display

Lb/kg/oz conversion

Semi-automatic (push-button) zero setting mechanism (SAZSM)

Semi-automatic (push-button) tare

RS-232C serial interface

Initial zero setting mechanism (IZSM)

Automatic zero setting mechanism (AZSM)

Piece counting: "The counting feature is not legal for trade" is labeled on the indicator face

Models HV-XXKGV/HV-XXKGL: Plastic indicator housing

Models with the V suffix have a fluorescent display, use AC line power and offer an optional built-in dot matrix printer

Models with the L suffix have a liquid crystal display and use six "D" size batteries or an AC adapter as a power source

Model HV-XXKV-WP: Stainless steel indicator housing

Models have a fluorescent display and use AC line power

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program (NTEP) and was found to comply with the applicable technical requirements of Handbook 44, "Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Effective Date: March 13, 2002



Ronald D. Murdock
Chairman, NCWM, Inc.



Louis E. Straub
Chairman, National Type Evaluation Program Committee

Issue date: March 14, 2002

Note: The National Conference on Weights and Measures does not "approve", "recommend", or "endorse" any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.